



#### LINEAR REGRESSION EQUATION

As ICP-MS, in µg/L = .107 + .954 \* As cICP-MS, in µg/L;  $R^2 = .998$

#### Confidence Intervals As ICP-MS, in µg/L vs. As cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	.107	-.091	.305
As cICP-MS, in µg/L	.954	.935	.972